



August 19, 2004

A-DEC INC.
ENGINEERING
TOM LOUISELL
2601 CRESTVIEW DRIVE
PO BOX 111
NEWBERG OR 97132

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: A-DEC INC.
Product Name: DENTAL UNIT
Model Number(s): CONTINENTAL UNITS: 2080, 2081, 2082, 2150, 2151, 2152, 2180, 2181 AND 2182
RADIUS UNITS: 2122 AND 2132; EXCELLENCE UNITS: 2070, 2071 AND 2072
Product File No: 20040429

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters Comm 82 through 84, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of January 2010.

This approval is contingent upon compliance with the following stipulation(s):

- A reduced pressure principle backflow preventer (ASSE 1013 or CAN/CSA B64.4) or air-gap must be installed on the water supply pipes serving this product.

This approval letter shall be incorporated with your previously approved plans and/or specifications approved under product file number 19990427.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Michael J. Beckwith, CIPE
Plumbing Product Reviewer
phone: 608-266-6742
fax: 608-267-9566
e-mail: mike.beckwith@wi.gov